



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

RPC-2.92 side

According to

Mechanically compatible with

IEC 61169-35

RPC-3.50 and SMA

Mini-SMP side

According to

Mechanically compatible with

MIL-STD-348A, Fig.328

GPPO™ (Gilbert Engineering Co., Inc.)
and SSMP™ (Connectors Devices, Inc.)

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts

- Center conductor
- Outer conductor RPC-2.92
- Outer conductor Mini-SMP
- Coupling Nut
- Dielectric

Material

- CuBe
- Stainless steel
- CuBe
- Stainless steel
- PS

Plating

- Gold, min. 1.27 µm, over nickel
- Passivated
- Gold, min. 1.27 µm, over nickel
- Passivated

Electrical data

Frequency	DC to 40 GHz
Return loss	≥ 28 dB, DC to 18 GHz ≥ 20 dB, 18 GHz to 40 GHz

Mechanical data

	RPC-2.92	Mini-SMP
Mating cycles	≥ 500	≥ 100
Maximum torque	1.70 Nm	
Recommended torque	0.90 Nm	
Engagement force		Full detent 19 N typical
Disengagement force		Full detent 29 N typical
Gauge	0.00 mm to 0.08 mm	0.00 mm to 0.08 mm

General standard definition

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset Z_o / Impedance / Z_o	50 Ω
Offset Delay	41.4475 ps
Length (electrical) / Offset Length	12.43 mm
Offset Loss	3.70 G Ω /s
Loss	0.0133 dB/ $\sqrt{\text{GHz}}$

Environmental data

Operating temperature range ¹	+20 °C to +26 °C
Rated temperature range of use ²	0 °C to +50 °C
Storage temperature range	- 40 °C to +85 °C

RoHS compliant

¹ Temperature range over which these specification are valid.

² This range is underneath and above the operating temperature range, within the calibration adaptor is fully functional and could be used without damage.

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Packing

Standard 1 pce in box
Weight 3.6 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Marcel Panicke	03.11.09	Markus Müller	03.11.16	d00	16-1390	Marion Striegler	03.11.16

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de				Tel. : +49 8684 18-0 Email : info@rosenberger.de			Page 3 / 3
--	--	--	--	---	--	--	---------------